


NE<sup>®</sup>



Code: NE029

## NG PISTIM

EEG & tCS

The next generation  
of the hybrid tCS &  
EEG electrodes

NE  
neuroelectrics<sup>®</sup>

### DESCRIPTION

The NG Pistim consists of two pieces: the fastener (top part) and the threaded washer (bottom part). The fastener is based on a Ag/AgCl sintered pellet with a 12 mm diameter, providing a  $\pi \text{ cm}^2$  circular contact area.

### MAINTENANCE


After each use, clean the NG Pistim using a tissue and water. Avoid direct sunlight exposure and contact with metals when the electrode is not being used.

### DURABILITY


After 10 hours of stimulation, the Ag/AgCl is expected to be consumed and the electrode needs to be replaced. Nevertheless, the time of Ag/AgCl consumption depends on the type and intensity of the stimulation current.

## USE INSTRUCTIONS


**1** – Insert the threaded washer in the desired position of the neoprene cap, and put the cap on.



**2** – Inject conductive gel inside the threaded washer with the syringe.



**3** – Screw the fastener onto the washer.



**4** – Connect the cable to the electrode connector.

